



**Republic of the Philippines**  
**WESTERN MINDANAO STATE UNIVERSITY**  
**Zamboanga City**  
**Telephone No.: 062-991-7875**  
**www.wmsu.edu.ph**

## INVITATION TO BID

The Western Mindanao State University, through its Bids and Awards Committee (BAC), is inviting PhilGEPS registered suppliers to apply for eligibility and to submit bids for the item mentioned hereunder:

Name of Project : **Supply and Delivery of Various Laboratory Equipment for the CSM-Research**

Approved Budget Cost : **Php 32,000.00**

Purchase Request No.: **20-05-096**

Closing Date: **July 21, 2020**

Description:

- 1.)      1    Unit      **Digital pH Meter**  
Specifications:
  - \* Description: single-junction
  - \* Type: sealed
  - \* pH Range: 0 to 14
  - \* Temperature Range: 0 to 80 degree Celsius
  - \* Length: 112 x 12 (mm)
  
- 2.)      1    Unit      **Digital Thermometer**  
Specifications:
  - \* Pen-style thermometer; with magnetic clip that allow thermometer to be attached to any metal surfaces
  - \* Range: -58 to 572 degree Fahrenheit (-50 to 300 degree Celsius)
  - \* Resolution: 0.1 degree Fahrenheit (1.0 degree Celsius)
  - \* Accuracy: +/-1.8 degree Fahrenheit (1.0 degree Celsius)
  - \* Display: 3 1/2 digit, 0.25;H
  - \* Dimensions: Head - 3.5"L x 1"W x 0.5"D, Probe - 3.25"L x 0.13: diameter
  
- 3.)      1    Piece      **Digital Salinity Meter**  
Specifications:
  - \* Range: 2ppt to 50ppt
  - \* Display: 3 1/2 digit LED
  - \* Dimension: at least 275 x 175 x 75mm
  - \* Resolution: 0.001ppt to 1ppt
  - \* Power: 220V, 60H

Features:

  - \* Micro Controller Based
  - \* Auto Ranging Facility
  - \* Auto Cell Constant Measurement
  
- 4.)      1    Piece      **Digital Light Meter**  
Specification:
  - \* Range: Lux-20/200/2000/20,000/200,000; Footcandles-

20/200/2000/20,000

\* Resolution: 0.01 lux/ftc

\* Accuracy: +/-13%

\* Output: none

\* Display: 3 1/2 digit LCD

\* Dimension: at least 6.3W x 13H x 3.8D cm

\* Power Source: 9V Battery

The criteria to be used for the eligibility check of the prospective bidders, examination and evaluation of bids, post-qualification and all matters relevant to this procurement shall be in accordance with Republic Act. No. 9184 (The Government Procurement Reform Act) and its Implementing Rules and Regulations.

Interested bidders are required to submit their valid and current Mayor's Permit, and PhilGEPS Registration, upon the submission of quotation.

Award of contract shall be made to the lowest calculated and responsive bid, which complies with the necessary description as stated above and other terms and conditions stated in the price quotation form.

Any interlineations, erasures or overwriting shall be valid only if they are signed or initiated by the bidder or his/her duly authorized representative/s.

Submission of Quotation and eligibility documents is on or before 10:00 AM of **July 21, 2020** at the BAC Office, Ground Floor Executive Building, Western Mindanao State University, Normal Road, Baliwasan, Zamboanga City. Open submission may be submitted manually, email ([bacsecretariate@wmsu.edu.ph](mailto:bacsecretariate@wmsu.edu.ph)) or through facsimile at (062)991-7875.

For inquiries, you may coordinate with the BAC Secretariat at telephone no. (062)991-7875.